



NVLAP LAB CODE 100396-0

• P.O. Box 489 • 1350 Tolland Road • Rollinsville, CO 80474 • Phone: (303) 258-0100 • FAX: (303) 258-0775 •
• www.criteriontech.com •

**EMC QUALIFICATION
TEST REPORT
ZYNEX MEDICAL, INC.
NEURO-MOVE, NM-900**

TESTED TO CONFORM WITH:

EMISSIONS STANDARDS **IMMUNITY STANDARDS**

FOR

INDUSTRIAL, SCIENTIFIC AND MEDICAL (ISM)

Test Report Number: 071010-1176

Date of Issue: February 13, 2008

Date of Test Completion: February 6, 2008

Manufacturer's Address: 8022 Southpark Circle, Unit 100
Littleton, CO 80120

Phone: 303-707-0203

Approved by:

A handwritten signature in blue ink, appearing to read "David Schmitt", written over a horizontal line.

Laboratory Director

CONFIDENTIAL AND PROPRIETARY

DISCLAIMERS

Report is the confidential property of the client. For the protection of our clients and ourselves, extracts from this test report cannot be produced without prior written approval from Criterion Technology. Reproduction of the complete report can be performed at the client's discretion.

Client is aware that Criterion Technology has performed testing in accordance with the applicable standard(s). Test data is accurate within ANSI parameters for emissions testing, unless a specific level of accuracy has been defined in writing prior to testing, by Criterion Technology and the client.

Criterion Technology reports apply only to the specific Equipment Under Test (EUT) sample(s) tested under the test conditions described in this report. If the manufacturer intends to use this report as a document demonstrating compliance of this model, additional models of this product must have electrical and mechanical characteristics identical to the device tested for this report. Criterion Technology shall have no liability for any deductions, inferences, or generalizations drawn by the client or others from Criterion Technology issued reports.

Criterion Technology liability is limited to the amount invoiced for the testing of this EUT and the contents of this report are not warranted.

Compliance with the appropriate governmental standards is the responsibility of the manufacturer.

Any questions regarding this report should be directed to:

Laboratory Director
Criterion Technology Corp.
P.O. Box 489
1350 Tolland Road
Rollinsville, Colorado 80474
Phone: (303) 258-0100
Fax: (303) 258-0775
mailto:laboratory_director@criteriontech.com

NVLAP Note: Criterion Technology is accredited by the National Voluntary Laboratory Accreditation Program (NVLAP) for the specific scope of accreditation under Lab Code 100396-0. Test methods included in Lab Code 100396-0 are

12/CIS11a - IEC/CISPR 11, edition 3.1 (1999-08)
12/CIS11b - IEC/CISPR 11 (2003) & EN 55011 (1998), A1 (1999), A2 (2002)
12/CIS11c - IEC/CISPR 11 (1997), A1 (1999), A2 (2002)
12/CIS11d - EN 55011 (1998), A1 (1999), A2 (2001)
12/CIS22 - IEC/CISPR 22 (1997) and EN 55022 (1998) + A1 (2000)
12/CIS22a - IEC/CISPR 22 (1993) and EN 55022 (1994), Amendment 1:1995 & Amendment 2: 1996
12/CIS22b - CNS13438 (1997)
12/EM02a - IEC 61000-3-2, Edition 2.1 (2001-10) and EN 61000-3-2 (2000), and AS/NZS 2279.1 (2000)
12/EM02b - IEC 61000-3-2, Second Edition (2000-08)
12/EM02c - BS EN 61000-3-2, ED. 2 (2001); IEC 61000-3-2, ED. 2 (2000)
12/EM03 - ICE 61000-3-3 (1995); EN 61000-3-3 (1995); AS/NZS 2279.3 (1995)
12/EM03a - ICE 1000-3-3 (1994-12)
12/EM03b - ICE 61000-3-3 Edition 1.1 (2002-03) & EN 61000-3-3, A1 (2001)
12/EM03c - ICE 61000-3-3 (1994) with Amendment 1 (2001)
12/EM03d - ICE 61000-3-3 (1995) + A1 (2001)
12/FCC15b - ANSI C63.4 (2003) with FCC Method 47 CFR Part 15, Subpart B
12/T51 - AS/NZS CISPR 22 (2002) and AS/NZS 3548 (1997)
12/101 - IEC 61000-4-2, Ed. 2.1 (2001), A1, A2; EN 61000-4-2
12/102 - IEC 61000-4-3, Ed. 2.0 (2002-03); EN 61000-4-3 (2002)
12/103 - IEC 61000-4-4 (1995), A1 (2002), A2 (2001); EN 61000-4-4
12/104 - IEC 61000-4-5, Ed. 1.1 (2001-04); EN 61000-4-5
12/105 - IEC 61000-4-6, Ed. 2.0 (2003-05); EN 61000-4-6
12/106 - IEC 61000-4-8, Ed. 1.1 (2001); EN 61000-4-8
12/107 - IEC 61000-4-11, Ed. 1.1 (2001-03); EN 61000-4-11

The NVLAP Logo on the front cover of this report applies only to data taken for the above test methods.

This report may contain data which is not covered by the NVLAP accreditation.

This report must not be used by the client to claim product endorsement by NVLAP or any agency of the U.S. Government.

Criterion Technology has been accredited by the following groups: NVLAP(#100396-0), FCC(#90688), BSMI(#SL2-IN-007R), VCCI(#1255) 3&10 meter site (#R-2097), Community Shield room(#C-2269), Open Area Site(#C-2270), Nemko(#ELA-214), NMI (EU Competent Body Accreditation) and Industry Canada(#C 3301). The National Institute for Standards and Technology (NIST) has designated Criterion Technology a Conformity Assessment Body (CAB) for Taiwan (BSMI # SL2-IN-E-007R). Criterion Technology instrumentation and accessories used to test products for compliance to the indicated standards are calibrated regularly in accordance with ISO 9002, ISO 17025, ANSI/NC SL Z540-I-1994 and are traceable to national standards.